

Uncertainty (%)	MAYP11868 (210–350 nm)	USB2000 (250–400 nm/400–800 nm)
Lamp irradiance standard	3.16	3.00–2.40/2.40–1.60
Converting (irradiance to radiance)	1.27	1.27
Spectral radiometer	1.44	1.21
Total	3.70	3.48–3.00/3.00–2.38