

Uncertainty (%)	Diffuse plate system (210–350 nm)	Integrating sphere (250–400 nm/400–800 nm)
Surface light source	< 2.00	< 2.00
Spectral radiometer	3.70	3.48–3.00/3.00–2.38
Total	< 4.21	< 4.02–3.61/3.61–3.11